

Certificate



Certificate for the Interlaboratory Comparison of Olfactometry 2011

This certificate was issued to confirm the participation of the laboratory with the coding number

C2 (decoding by coding body IFEP GmbH)

in the international interlaboratory comparison olfactometry, held in June 2011. The aim of the test was to determine the precision r and accuracy A_{od} according to EN 13725:2003. Results of the participant are as follows:

quality parameter assessed using 1-butanol in nitrogen	participation (requirements according to EN 13725)
$A_{od} = 0,099$	complies with requirement: $A \leq 0,217$
$r = 0,409$	complies with requirement: $r \leq 0,477$

The odorant used was: 1-butanol in nitrogen
Number of participating laboratories: 35

Kiel, August 24, 2011

(Björn Maxeiner - director, head of inspection body, OLFAtec)